

AIHA PAT Programs' Proficiency Analytical Testing (BePAT) Program

(administered by RTI International)

SGS Galson Laboratories, Inc.

Participant ID: 100324

Round 40

April 15, 2016

Sample ID	Results Qualifier	Reported Results (μg)	Round Mean (μg)	Expected (μg)	Recovery (%)	Number of Participants	Standard Deviation	Coefficient of Variation	Rating	Comments
1		15.811	15.8	16.1	98.2	35	1.35	0.0854	A	
2		2.1525	2.05	2.11	102	35	0.149	0.0727	A	
3		21.567	20.8	22.1	97.6	35	2.19	0.105	A	
4		0.768	0.741	0.848	90.6	35	0.0847	0.114	**	Not Graded
blank	<	0.015								

Rating Key

A - Acceptable (between 75% and 125% of Expected Value)

L - < 75% of Expected Value

H - > 125% of Expected Value

* - No rating assigned

NR - Not reported

Round Status: pass

Program Status: P

P - Proficient

NP - Not Proficient

I - Indeterminate

** Sample 4 was determined to be unsuitable for the proficiency test and was not graded. For BePAT round 40, proficiency is achieved by passing at least two of the three remaining samples. See details in the Final Report to Participants.

Your temporary laboratory ID is #19 for this round.

This Temp ID is used on the Final Report to Participants to identify your individual results.